



Medprotex BV

Gaetano Martinolaan 63
6229 GD, Maastricht
Nederland

Your notice of	Your reference	Date
15-06-2020		07-10-2020

Analysis Report 2020-10

Required tests

EN 14683 (2019) + AC (2019) EN 14683 (2019) - annex B (2019) Bacterial filtration efficiency

SOP used

AMIBM SOP 2.0;
Date 29.6.2020

Identification number	Information given by the client	Date of receipt
	Medprotex Light Non-Medical	15-06-2020

A handwritten signature in blue ink, appearing to read 'Luisa Bortesi'.

Luisa Bortesi

Disclaimer: This report may not be reproduced without written permission of AMIBM. The results of the analysis cover the received samples. AMIBM is not responsible for the representativeness of the samples. In assessing compliance with the specifications, we did not take into account the uncertainty on the test results. AMIBM has applied for accreditation at the RvA.

Visiting address
Urmonderbaan 22
6167 RD Geleen

Postal address
P.O. Box 616
6200 MD Maastricht
The Netherlands

T +31 (0)43 388 22 22
F +31 (0)43 388 52 47

Bank account: 065.76.18.705
IBAN: NL05 INGB 0657 6187 05
BIC: INGBNL2A
VAT identifier EU
NL0034.75.268.B01

www.maastrichtuniversity.nl
Masktest-amibm@maastrichtuniversity.nl
KvK nr.: 50169181

Reference: Medprotex Light Non-Medical

Bacterial filtration efficiency

Date of ending the test	08.07.2020
Standard used	EN 14683 (2019) - annex B (2019)
Product standard	EN 14683 (2019) + AC (2019)
SOP version used	AMIBM SOP 2.0; Date 29.6.2020
Number of tested masks	5
BFE area tested	±49 cm ²
Mask conditioning	≥4 h at 21 ± 5° C and 85 ± 5% RH
Mask/bacteria contact side	Inner side
Challenge bacterial strain used	<i>Staphylococcus aureus</i> ATCC6538
Bacterial challenge per test	1700-3000 CFU
Total test time	1 min. challenge delivery + 1 min. without challenge (continuous air flow)
Flow rate	28.3 l/min
Positive control	Tests performed without filter in air stream
Negative control	Test performed without bacterial challenge
Deviation from the standard	

Results

B = Bacterial filtration efficiency (%)

C = Mean of the total plate counts for the positive control runs

T = Total count for the tested mask specimen

$$B = \frac{(C - T)}{C} \times 100$$

Mask #	B (%)
1	>99.9*
2	>99.9*
3	>99.9*
4	>99.9*
5	>99.9*

*no detected colony on any TSA sampler plates

Mean particle size of the bacterial aerosol challenge: 3.0 µm

Controls

Mean positive controls 1835 CFU
Negative control < 1 CFU

This test report is valid for products used in relation to the current Covid-19 health crisis and for products which are not entering the regular distribution channels. Cfr Commission Recommendation (EU) 2020/403 of 13 March 2020 on conformity assessment and market surveillance procedures within the context of the COVID-19 threat